

INSTRUCTIONS C24 ARYLOXY ACID PESTICIDES IN WATER

1.0 Sample Reception

- 1.1 Check that all the samples for which you are registered are accounted for. All breakages and shortages must be reported within 24 hours of sample receipt.
- 1.2 Store samples at $4\pm2^{\circ}$ C upon receipt. Samples are preserved to a pH < 2 with H₂SO₄. Extract within 21 days of receipt.
- 1.3 Check that all the parameters for which you are registered are correctly identified on the web data entry report form.
- 1.4 Inquiries regarding samples and their shipment may be directed to:

Mr. Guillaume Bourque Centre d'expertise en analyse environnementale du Québec Tel. 418 643-1301 ext. 352 Fax. 418 646-7612 email: guillaume.bourque@mddelcc.gouv.qc.ca

cc: Erinn Knight, CALA Program Administrator

fax: 613-233-5501 email: eknight@cala.ca

Inquiries must be made by facsimile or email only. Use the Nonconformance Form (see reverse) when sending a fax. Please include your CALA membership number on all correspondence.

2.0 Sample Analysis

- 2.1 Refer to P02-04-*CALA Program Description PT Catalogue* for approximate concentration range.
- 2.2 Proceed with testing using the routine analytical method identified in your recent application to the CALA program.

3.0 Reporting Results

- 3.1 Results must be reported by midnight of the study deadline (see the General Proficiency Testing Information sheet for details).
- 3.2 Report RDL (optional) if you want RDL accounted for in z scores.

4.0 Safety

4.1 The PT samples are designed for use by laboratory professionals familiar with environmental samples and potentially hazardous materials.

Rev 1.4 Page 1 of 2

PT SAMPLE NON-CONFORMANCE FORM

Attn: PT non-conformances	Study Number:
ENSURE THAT SAMPLES RECEIVED MATCH REPORT FORMS 1 - Laboratory Information	
Contact Name:	
<u>Laboratory Name</u>	
<u>Laboratory Address</u>	
Contact Telephone #	
Contact Facsimile #	
Contact e-mail:	
2 - Sample Details	
Date & Time of Arrival(YYYY,MM,DD,HH:MM):	
Tracking Number:	
Test Groups Received (e.g. C1, C2 etc.):	
Number of Boxes:	
3 - Description of Nonconformance	
4 - Requested Action	
5 - PT Provider Notes	

Rev 1.4 Page 2 of 2