

INSTRUCTIONS C31A BTEX AND PHCs IN SOIL

1.0 Sample Reception

- 1.1 All breakages and shortages must be reported within 24 hours of sample receipt.
- 1.2 Store samples at $4^{\circ}C + 3^{\circ}C$.
- 1.3 Check that all the parameters for which you are registered are correctly identified on the web data entry report page.
- 1.4 Inquiries regarding samples and their shipment may be directed to:

PT Non-conformances Information and Quality Management Environment and Climate Change Canada

fax: 905-336-8914

email: ec.ptnc.ec@canada.ca

cc: CALA Program Administrator

cc: Ken Middlebrook, CALA PT Manager

fax: 613-233-5501

email: programadmin@cala.ca email: kmiddlebrook@cala.ca

Inquiries must be made by facsimile or email only. Use the Nonconformance Form (see reverse) when sending a fax. Please include your CALA membership number on all correspondence.

2.0 Sample Analysis

- 2.1 Perform the analysis on these samples assuming that each sample contains 8 g dry weight of soil and 20 mL of methanol.
- 2.2 Refer to P02-04-*CALA Program Description PT Catalogue* for approximate concentration range.
- 2.3 Proceed with testing using the CALA registered method.
- 2.4 Results for F1 must be reported in the CCME analytical fraction.

3.0 Reporting Results

- 3.1 Results must be reported by midnight of the study deadline (see the General Proficiency Testing Information sheet for details).
- 3.2 Report results on a dry weight basis. Do not subtract BTEX or any PAHs from F1 fraction.

4.0 Safety

4.1 The PT samples are designed for use by laboratory professionals familiar with environmental samples and potentially hazardous materials.

Rev 1.3 Page 1 of 2

PT SAMPLE NON-CONFORMANCE FORM

Attn: PT non-conformances	Study Number:
ENSURE THAT SAMPLES RECEIVED MATCH REPORT FORMS	
1 - Laboratory Information	
Contact Name:	
<u>Laboratory Name</u>	
<u>Laboratory Address</u>	
Contact Telephone #	
Contact Facsimile #	
Contact e-mail:	
John Grand	
2 - Sample Details	
Date & Time of Arrival(YYYY,MM,DD,HH:MM):	
<u>Tracking Number:</u>	
Test Groups Received (e.g. C1, C2 etc.):	
Number of Boxes:	
3 - Description of Nonconformance	
4 - Requested Action	
5 - PT Provider Notes	

Rev 1.3 Page 2 of 2